


Product no **AS04 042S****PsaC | PSI positive control/quantitation standard****Product information**

<b>Format</b>	Lyophilized in glycerol.
<b>Quantity</b>	100 µl
<b>Reconstitution</b>	For reconstitution add 95 µl of sterile water. Note that due to glycerol in buffer, the lyophilized product appears as a dense liquid rather than a powder. Allow the product several minutes to solubilize after adding water. Mix thoroughly but gently. Avoid vigorous vortexing, as buffer contains detergent. Upon reconstitution, this standard is ready-to-load and does not require any additions or heating. See additional <b>Handling Instructions</b> below.  PsaC standard protein concentration: 0.10 pmol/µl.
<b>Storage</b>	Store lyophilized/reconstituted at -20°C; once reconstituted make aliquots to avoid repeated freeze-thaw cycles. Please remember to spin the tubes briefly prior to opening them to avoid any losses that might occur from material adhering to the cap or sides of the tube.
<b>Additional information</b>	<b>Handling Instructions</b>  *IMPORTANT: In our experience, viscous liquids are surprisingly stable; insufficient mixing is the most common reason for unsatisfactory results. Following mixing, briefly pulse in a microcentrifuge to collect material from cap.  Standard needs to be fully thawed and thoroughly mixed before each use. Proteins tend to stratify with the more dense layer after freezing. We recommend bringing the product to room temperature and either mixing by inverting or flicking tube 5-10 times. Pipetting up and down may also provide sufficient mixing, provided the tip is moved within the tube while taking up and expelling the liquid.

**Application information**

<b>Recommended dilution</b>	<b>Positive control:</b> a 2 µL load per well is optimal for most chemiluminescent detection systems.  <b>Standard curve:</b> 3 loads are recommended (eg. 0.5, 2 and 4 µL). For most applications a sample load of 0.2 µg of chlorophyll will give a PsaC signal in this range. Exact loads can vary with the sensitivity of your system and the abundance of the target protein in your samples.  <b>Note:</b> Optimal quantitation is achieved using moderate sample loads/well, generally 1 to 5 ug total protein.  A trial experiment may be required i) to bring your sample load within the standard curve range and ii) to obtain a signal that is strong enough to reliably quantify but not so strong as to consume ECL reagents too quickly or saturate your detection system. These goals may be achieved by adjusting both sample and standard loads.
<b>Expected   apparent MW</b>	11.5 kDa (larger than native protein due to the addition of His-tag), In most gels PsaC migrates between 9 and 14 kDa
<b>Additional information</b>	<b>Protein standard buffer composition:</b> Protein standard buffer composition: Glycerol 10%, Tris Base 141 mM, Tris HCl 106 mM, LDS 2%, EDTA 0.51 mM, SERVA® Blue G250 0.22 mM, Phenol Red 0.175 mM, pH 8.5, 0.1mg/ml PefaBloc protease inhibitor (Roche), 50mM DTT.
<b>Selected references</b>	<a href="#">Pipitone</a> et al. (2021). A multifaceted analysis reveals two distinct phases of chloroplast biogenesis during de-etiolation in Arabidopsis. <i>Elife</i> . 2021 Feb 25;10:e62709. doi: 10.7554/eLife.62709. PMID: 33629953; PMCID: PMC7906606. <a href="#">Rogowski</a> et al. (2021) Light as a substrate: migration of LHCII antennas in extended Michaelis-Menten model for PSI kinetics. <i>J Photochem Photobiol B</i> . 2021 Dec;225:112336. doi: 10.1016/j.jphotobiol.2021.112336. Epub 2021 Oct 19. PMID: 34736069. <a href="#">Levitan</a> et al. (2019). Structural and functional analyses of photosystem II in the marine diatom <i>Phaeodactylum tricorutum</i> . <i>Proc Natl Acad Sci U S A</i> . 2019 Aug 27;116(35):17316-17322. doi: 10.1073/pnas.1906726116. <a href="#">Lj</a> et al. (2016). A Hard Day's Night: Diatoms Continue Recycling Photosystem II in the Dark. <i>Front. Mar. Sci.</i> , 08 November 2016   <a href="http://dx.doi.org/10.3389/fmars.2016.00218">http://dx.doi.org/10.3389/fmars.2016.00218</a> <a href="#">Vandenhecke</a> et al. (2015). Changes in the Rubisco to photosystem ratio dominates photoacclimation across phytoplankton taxa. <i>Photosynth Res</i> . 2015 Apr 11.

## Application example



**Total protein** from *Trichodesmium* sp. (1) and *Thalassiosira* sp. (2). Recombinant PsaC protein standard (AS04 042S) (3-6) loaded at 0.5 pmoles, 0.3 0.1 and 0.05 pmoles. Molecular weight markers (MagicMark XP, Invitrogen) (7). Samples were separated on 4-12% NuPage (Invitrogen) **LDS-PAGE** and blotted 1h to **PVDF**. Blots were blocked immediately following transfer in 2% blocking reagent in 20 mM Tris, 137 mM sodium chloride pH 7.6 with 0.1% (v/v) Tween-20 (TBS-T) for 1h at room temperature with agitation. Blots were incubated in the primary antibody at a dilution of 1: 50 000 for 1h at room temperature with agitation. The antibody solution was decanted and the blot was rinsed briefly twice, then washed once for 15 min and 3 times for 5 min in TBS-T at room temperature with agitation. Blots were incubated in secondary antibody (anti-hen IgY horse radish peroxidase conjugated) diluted to 1:50 000 in 2% blocking solution for 1h at room temperature with agitation. The blots were washed as above and developed for 5 min with chemiluminescence detection reagent according the manufacturers instructions. Images of the blots were obtained using a CCD imager (FluorSMax, Bio-Rad) and Quantity One software (Bio-Rad).

**Note:** Optimal quantitation is achieved using moderate sample loads per gel lane, generally 0.5 to 2.5 ug total protein, depending on the abundance of the target protein.

**Quantitation:** When quantitated standards are included on the blot, the samples can be quantitated using the available software. Excellent quantitation can be obtained with images captured on the Bio-Rad Fluor-S-Max or equivalent instrument using Bio-Rad QuantityOne software. The contour tool is used to select the area for quantitation and the values are background subtracted to give an adjusted volume in counts for each standard and sample. Using above protocol linear standard curves are generated over 1-1.5 orders of magnitude range in target load. It is important to note that immunodetections usually show a strongly sigmoidal signal to load response curve, with a region of trace detection of low loads, a pseudolinear range and a region of saturated response with high loads. For immunoquantitation it is critical that the target proteins in the samples and the standard curve fall within the pseudolinear range. Our total detection range using this protocol spans over 2 orders of magnitude, but the quantifiable range is narrower.

Quantitative western blot: [detailed method description](#).