


Product no **AS16 3155****Anti-ADK | Adenylate kinase****Product information**

Immunogen	Protein derived from <i>Escherichia coli</i> , Uniprot: P69441
Host	Rabbit
Clonality	Polyclonal
Purity	Serum
Format	Lyophilized
Quantity	50 µl
Reconstitution	For reconstitution add 50 µl, of sterile water
Storage	Store lyophilized/reconstituted at -20°C; once reconstituted make aliquots to avoid repeated freeze-thaw cycles. Please remember to spin the tubes briefly prior to opening them to avoid any losses that might occur from material adhering to the cap or sides of the tube.

Application information

Recommended dilution	1: 100 000 (WB)
Expected apparent MW	23.5 kDa (<i>E.coli</i>), 24.3 kDa (<i>S.cerevisiae</i>)
Confirmed reactivity	<i>Escherichia coli</i>
Predicted reactivity	
Not reactive in	No confirmed exceptions from predicted reactivity are currently known
Selected references	Tükenmez et al. (2016) . Linkage between Fitness of Yeast Cells and Adenylate Kinase Catalysis. PLoS One. 2016 Sep 19;11(9):e0163115. doi: 10.1371/journal.pone.0163115. eCollection 2016.

Application example

Protein samples were prepared using TCA-extraction method. *S. cerevisiae* *adk1Δ* strains expressing either *S. cerevisiae* Adk1 protein (24.3 kDa) or *E. coli* adk protein (23.5 kDa) were cultivated logarithmically, and 5 OD-units of cells were collected for TCA-extraction method. In the end, protein pellets were re-suspended with 300 µl of TCA-Laemmli Loading Buffer and denatured at 100°C for 10 minutes. Total protein samples (8 µl of each sample) were separated by 12% SDS-PAGE and transferred to nitrocellulose membranes using a wet transfer protocol. The membranes were blocked with 5% dry-milk in 1x TBST for 1.5 hours at room temperature (RT) with agitation and then they were incubated with primary antibody (anti-adk antibody, 1:2,000) in 5% dry-milk in 1x TBST overnight at 4°C with agitation. The primary antibody solution was decanted and the membranes were washed 3 times for 10 minutes with 1x TBST with agitation. The membranes were then incubated with secondary antibody (HRP conjugated anti-rabbit IgG, 1:4,000) in 5% dry-milk in 1x TBST for 1 hour at RT with agitation. The membranes were washed as above and the signals were detected using chemiluminescent detection reagent and exposing the membranes in LAS4000 machine (GE-healthcare) for 10 seconds. Followed-up experiments revealed that lowering the primary antibody concentration from 1:2,000 to 1:50,000-100,000 reduces the unspecific background signals and lowers the risk of target signal oversaturation.

Courtesy of Dr. Hasan Tukenmez, Umeå University, Sweden